



NIGHT 14		Theme: starting and ending with closed dome: cloudy and strong winds.		
Night 19/20 June (171/xx) Gijs				
Seq	Local time	Activity	Comment	
		No opening of dome due to strong winds.		
<b>FLAT171 1-55</b>	20:20	Domeflats u,g,r,l,z,Halpha,strV,ugri	Ugri at 2 exp times: 100s for u and 0.8 for quadrants	
<b>BIAS171 1-4</b>	21:45	bias	Makes calibration unit to turn off.	
<b>FLAT171 56-63</b>	22:00	<i>Domeflats G,r,g,r</i>	Filterwheel re-insertion test	
<b>BIAS171 5-19</b>	22:15	<i>Bias</i>	No filterwheel moving	
<b>Bias171 20-29</b>	22:30	Bias	With filterwheel moving: testing EMC	
	~24:00	Interlock from maximum torque reached after rotating rotator 2deg . Corotator was in standby. Windspeed too high to allow inspection in the dome		
<b>FLAT171 64</b>	02:15	Flat to check detector is in order	Seems in order	
<b>BIAS171</b>	02:20	bias	Template invoked to turn off calibration unit	

